

STANDARD F**H. pylori Ag FIA**  **SD BIOSENSOR**

STANDARD F H.pylori Ag FIA is a fluorescent immunoassay for the detection of H.pylori antigen in human stool samples.



| Specification | |
|-------------------|------------------|
| Specimen Type | Feces |
| Specimen Volume | 40-70 mg |
| Testing Time | 10 mins. |
| Storage Condition | 2-30°C / 36-86°F |

TEST PERFORMANCE

| | Sensitivity | Specificity |
|----------------|-------------|----------------|
| Internal Study | 100% (5/5) | 100% (150/150) |

ORDERING INFORMATION

| Category | Product | Pack Size |
|------------------|-----------------------------|-----------|
| Gastrointestinal | STANDARD F H. pylori Ag FIA | 25 Test |

STANDARD F**C. difficile GDH FIA**  **SD BIOSENSOR**

STANDARD F C. difficile GDH FIA qualitatively analyzes Clostridium difficile Glutamate Dehydrogenase (GDH) antigen in human stool specimen using Fluorescent Immunoassay (FIA). With a simple test method, the test result can be obtained within 15 minutes with its high sensitivity and specificity.



| Specification | |
|-------------------|------------------|
| Specimen Type | Feces |
| Specimen Volume | 40-70 mg |
| Testing Time | 15 mins. |
| Storage Condition | 2-30°C / 36-86°F |

TEST PERFORMANCE

| | Sensitivity | Specificity |
|---------------------|----------------|----------------|
| Culture GDH | 95.24% (80/84) | 100% (150/150) |
| EIA GDH (Well type) | 100% (77/77) | 100% (79/79) |

ORDERING INFORMATION

| Category | Product | Pack Size |
|------------------|--------------------|-----------|
| Gastrointestinal | STANDARD F GDH FIA | 25 Test |

STANDARD F**C. difficile Toxin FIA**  **SD BIOSENSOR**

STANDARD F C. difficile Toxin A/B FIA qualitatively analyzes Clostridium difficile Toxin A/B in human stool specimen using Fluorescent Immunoassay (FIA). With a simple test method, the test result can be obtained within 15 minutes with its high sensitivity and specificity.



| Specification | |
|-------------------|------------------|
| Specimen Type | Feces |
| Specimen Volume | 40-70 mg |
| Testing Time | 15 mins. |
| Storage Condition | 2-30°C / 36-86°F |

ORDERING INFORMATION

| Category | Product | Pack Size |
|------------------|-----------------------------------|-----------|
| Gastrointestinal | STANDARD F C. difficile Toxin FIA | 25 Test |



TruLaboratories
Corporation

+632 82-TRULAB (878522)
+632 8287.3683
+63917.878.2288 • +63925.878.2288
www.trulaboratories.com